McCampbell Analytical, Inc.

"When Quality Counts"

Analytical Report

WorkOrder:	2404P75
Report Created for:	Gibco Environmental Services 2163 Meeker Ave STE. 195 Richmond, CA 94804
Project Contact:	Walloh Gibson
Project P.O.:	
Project:	Edna Brewer Middle Sch. Drinking Water Test
Project Location:	3748 13rh Ave Oakland CA 94610
Project Received:	04/30/2024

Analytical Report reviewed & approved for release on 05/07/2024 by:

Yen Cao
Project Manager

The report shall not be reproduced except in full, without the written approval of the laboratory. The analytical results relate only to the items tested. Results reported conform to the most current NELAP standards, where applicable, unless otherwise stated in a case narrative.

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Glossary of Terms & Qualifier Definitions

Client: Gibco Environmental Services **WorkOrder:** 2404P75
Project: Edna Brewer Middle Sch. Drinking Water Test

Glossary Abbreviation

%D	Serial Dilution Percent Difference
95% Interval	95% Confident Interval
CCV	Continuing Calibration Verification.
CCV REC (%)	% recovery of Continuing Calibration Verification.
CPT	Consumer Product Testing not NELAP Accredited
DF	Dilution Factor
DI WET	(DISTLC) Waste Extraction Test using DI water
DISS	Dissolved (direct analysis of 0.45 µm filtered and acidified water sample)
DLT	Dilution Test (Serial Dilution)
DUP	Duplicate
EDL	Estimated Detection Limit
ERS	External reference sample. Second source calibration verification.
ITEF	International Toxicity Equivalence Factor
LCS	Laboratory Control Sample
LCS2	Second LCS for the batch. Spike level is lower than that for the first LCS; applicable to method 1633.
LQL	Lowest Quantitation Level
MB	Method Blank
MB % Rec	% Recovery of Surrogate in Method Blank, if applicable
MDL	Method Detection Limit [1]
ML	Minimum Level of Quantitation
MS	Matrix Spike
MSD	Matrix Spike Duplicate
NA	Not Applicable
ND	Not detected at or above the indicated MDL or RL
NR	Data Not Reported due to matrix interference or insufficient sample amount.
PDS	Post Digestion Spike
PF	Prep Factor
RD	Relative Difference
RL	Reporting Limit [2]
RPD	Relative Percent Difference
RRT	Relative Retention Time
RSD	Relative Standard Deviation
SNR	Surrogate is diluted out of the calibration range
SPK Val	Spike Value

[1] MDL is the minimum measured concentration of a substance that can be reported with 99% confidence that the measured concentration is distinguishable from method blank results. Definition and Procedure for the Determination of the Method Detection Limit, Revision 2, 40CFR, Part 136, Appendix B, EPA 821-R-16-006, December 2016. Values are based upon our default extraction volume/amount and are subject to change.

[2] RL is the lowest level that can be reliably determined within specified limits of precision and accuracy during routine laboratory operating conditions. (The RL cannot be lower than the lowest calibration standard used in the initial calibration of the instrument and must be greater than the MDL.) Values are based upon our default extraction volume/amount and are subject to change.

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Glossary of Terms & Qualifier Definitions

Client: Gibco Environmental Services **WorkOrder:** 2404P75

Project: Edna Brewer Middle Sch. Drinking Water Test

SPKRef Val	Spike Reference Value
SPLP	Synthetic Precipitation Leachate Procedure
ST	Sorbent Tube
TCLP	Toxicity Characteristic Leachate Procedure
TEQ	Toxicity Equivalents
TNTC	“Too Numerous to Count;” greater than 250 colonies observed on the plate.
TZA	TimeZone Net Adjustment for sample collected outside of MAI's Coordinated Universal Time (UTC). (Adjustment for Daylight Saving is not accounted.)
WET (STLC)	Waste Extraction Test (Soluble Threshold Limit Concentration)

Quality Control Qualifiers

F13	Indigenous sample results too high for a representative matrix spike analysis.

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-1A/Flow Water-Hall Opp Counsellor Off	2404P75-001A	Water	04/27/2024 13:45	ICP-MS6 123SMPL.d	292928

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 12:58

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-2B/Faucet-Sink Attend. Office	2404P75-002A	Water	04/27/2024 13:49	ICP-MS6 124SMPL.d	292928

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.5**	0.50	1	05/02/2024 13:01

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-3C/Drinking Fountain Opp Rm 24	2404P75-003A	Water	04/27/2024 13:53	ICP-MS6 125SMPL.d	292928

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.3**	0.50	1	05/02/2024 13:04

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-4D/Drinking Fountain-Opp. Rm 20	2404P75-004A	Water	04/27/2024 13:57	ICP-MS6 126SMPL.d	292928

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 13:07

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-5E/Drinking Fountain-Hall Adj. Cafeteria	2404P75-005A	Water	04/27/2024 14:00	ICP-MS6 129SMPL.d	292928

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 13:16

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-6F/Faucet-Sink-Staff Lunch Room	2404P75-006A	Water	04/27/2024 14:03	ICP-MS6 130SMPL.d	292928

Analytes	Result	RL	DF	Date Analyzed
Lead	**0.79**	0.50	1	05/02/2024 13:19

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-7G/Kitchen Prep Sink-Faucet-Sink	2404P75-007A	Water	04/27/2024 14:07	ICP-MS6 098SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**540**	5.0	10	05/02/2024 11:43

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-8G Flush/Kitchen Prep Sink-Faucet-Sink	2404P75-008A	Water	04/27/2024 14:08	ICP-MS6 132SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**8.1**	0.50	1	05/02/2024 13:25

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-9H/Drinking Fountain-Hall adj. RM 30	2404P75-009A	Water	04/27/2024 14:16	ICP-MS6 133SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.4**	0.50	1	05/02/2024 13:28

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-10I/Flow Water Adj. Off Bldg. 30	2404P75-010A	Water	04/27/2024 14:20	ICP-MS6 134SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.0**	0.50	1	05/02/2024 13:31

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-11J/Flow Water Adj. Off Bldg. 30	2404P75-011A	Water	04/27/2024 14:21	ICP-MS6 135SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 13:34

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-12K/Drinking Fountain-Boys Locker RM	2404P75-012A	Water	04/27/2024 14:24	ICP-MS6 136SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**7.0**	0.50	1	05/02/2024 13:37

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-13L/Drinking Fountain-Girls Locker RM	2404P75-013A	Water	04/27/2024 14:29	ICP-MS6 137SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**4.0**	0.50	1	05/02/2024 13:40

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-14M/Drinking Fountain-Girls Locker RM	2404P75-014A	Water	04/27/2024 14:32	ICP-MS6 138SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**8.5**	0.50	1	05/02/2024 13:43

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-15N/Exterior Drinking Fountain Gym	2404P75-015A	Water	04/27/2024 14:35	ICP-MS6 141SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 13:52

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-16O/Exterior Drinking Fountain-Gym	2404P75-016A	Water	04/27/2024 14:37	ICP-MS6 142SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 13:55

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-17P/Drinking Fountain-Boys Locker Roo	2404P75-017A	Water	04/27/2024 14:39	ICP-MS6 143SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**2.0**	0.50	1	05/02/2024 13:58

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-18Q/Drinking Fountain-Hall Adj. RM 41	2404P75-018A	Water	04/27/2024 14:45	ICP-MS6 144SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**0.67**	0.50	1	05/02/2024 14:01

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-19R/Fauect-Sink-RM 42-Bldg 40	2404P75-019A	Water	04/27/2024 14:48	ICP-MS6 145SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**2.4**	0.50	1	05/02/2024 14:04

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-20S/Fauect-Sink-RM 43	2404P75-020A	Water	04/27/2024 14:52	ICP-MS6 147SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**270**	0.50	1	05/02/2024 14:10

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-21T/Faucet-Sink-RM 43	2404P75-021A	Water	04/27/2024 14:57	ICP-MS6 148SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.5**	0.50	1	05/02/2024 14:13

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-22U/Faucet-Sink-RM 43	2404P75-022A	Water	04/27/2024 14:59	ICP-MS6 149SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.0**	0.50	1	05/02/2024 14:16

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-23V/Faucet-Sink-RM 43	2404P75-023A	Water	04/27/2024 15:02	ICP-MS6 150SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**0.88**	0.50	1	05/02/2024 14:19

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-33W/Drinking Fountain-RM 50	2404P75-024A	Water	04/27/2024 15:08	ICP-MS6 153SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**4.9**	0.50	1	05/02/2024 14:28

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-34X/Faucet-Sink-Bldg 50	2404P75-025A	Water	04/27/2024 15:11	ICP-MS6 154SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**4.9**	0.50	1	05/02/2024 14:31

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-35Y/Faucet-Sink-Bldg 50	2404P75-026A	Water	04/27/2024 15:12	ICP-MS6 155SMPL.d	292933

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.8**	0.50	1	05/02/2024 14:34

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-36Z/Drinking Fountain-RM 55	2404P75-027A	Water	04/27/2024 15:17	ICP-MS6 108SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	**3.1**	0.50	1	05/02/2024 12:13

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-37AA/Faucet Sink-RM 55	2404P75-028A	Water	04/27/2024 15:19	ICP-MS6 156SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.6**	0.50	1	05/02/2024 14:37

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-38BB/Drinking Fountain-Rm 55	2404P75-029A	Water	04/27/2024 15:20	ICP-MS6 157SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	**180**	0.50	1	05/02/2024 14:40

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-39CC/Exterior Drinking Fountain-Main Bl	2404P75-030A	Water	04/27/2024 15:25	ICP-MS6 158SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 14:43

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-40DD/Exterior Drinking Fountain-Main Bl	2404P75-031A	Water	04/27/2024 15:26	ICP-MS6 159SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 14:46

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-41EE/Exterior Drinking Fountain-Main Bl	2404P75-032A	Water	04/27/2024 15:30	ICP-MS6 160SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 14:49

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-42FF/Exterior Drinking Fountain Bldg. 20	2404P75-033A	Water	04/27/2024 15:32	ICP-MS6 161SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 14:52

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-43GG/Exterior Drinking Fountain-Bldg 20	2404P75-034A	Water	04/27/2024 15:37	ICP-MS6 162SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	**0.66**	0.50	1	05/02/2024 14:55

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-44HH/Exterior Drinking Fountain-Bldg 20	2404P75-035A	Water	04/27/2024 15:39	ICP-MS6 165SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 15:04

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-45II/Exterior Drinking Fountain-Bldg 20	2404P75-036A	Water	04/27/2024 15:41	ICP-MS6 166SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.6**	0.50	1	05/02/2024 15:07

Analyst(s): WV

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Analytical Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Received:	04/30/2024 14:50	**Extraction Method:**	E200.8
Date Prepared:	05/02/2024	**Analytical Method:**	E200.8
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Unit:**	µg/L

Metals

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-46JJ/Exterior Drinking Fountain-Bldg 20	2404P75-037A	Water	04/27/2024 15:43	ICP-MS6 167SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 15:10

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-47KK/Flow Water Adj. Cafateria-Hallway-	2404P75-038A	Water	04/27/2024 15:47	ICP-MS6 168SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	**1.6**	0.50	1	05/02/2024 15:13

Analyst(s): WV

Client ID	Lab ID	Matrix	Date Collected	Instrument	Batch ID
042724-48LL/Flow Water Adj. Cafateria-Hallway-	2404P75-039A	Water	04/27/2024 15:48	ICP-MS6 169SMPL.d	292940

Analytes	Result	RL	DF	Date Analyzed
Lead	ND	0.50	1	05/02/2024 15:16

Analyst(s): WV

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Quality Control Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Prepared:	05/02/2024	**BatchID:**	292928
Date Analyzed:	05/02/2024	**Extraction Method:**	E200.8
Instrument:	ICP-MS6	**Analytical Method:**	E200.8
Matrix:	Drinking Water	**Unit:**	µg/L
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Sample ID:**	MB/LCS/LCSD-292928

QC Summary Report for Metals

Analyte	MB Result	MDL	RL			
Lead	ND	0.052	0.50	-	-	-

Analyte	LCS Result	LCSD Result	SPK Val	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Limit
Lead	51	52	50	102	104	85-115	1.84	20

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Quality Control Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Prepared:	05/02/2024	**BatchID:**	292933
Date Analyzed:	05/02/2024	**Extraction Method:**	E200.8
Instrument:	ICP-MS6	**Analytical Method:**	E200.8
Matrix:	Drinking Water	**Unit:**	µg/L
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Sample ID:**	MB/LCS/LCSD-292933 2404P75-007AMS/MSD

QC Summary Report for Metals

Analyte	MB Result	MDL	RL			
Lead	ND	0.052	0.50	-	-	-

Analyte	LCS Result	LCSD Result	SPK Val	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Limit
Lead	53	54	50	105	108	85-115	2.49	20

Analyte	MS DF	MS Result	MSD Result	SPK Val	SPKRef Val	MS %REC	MSD %REC	MS/MSD Limits	RPD	RPD Limit
Lead	10	540	560	50	537.2	15,F13	38,F13	85-115	2.10	20

(Cont.)

CA ELAP 1644

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Quality Control Report

Client:	Gibco Environmental Services	**WorkOrder:**	2404P75
Date Prepared:	05/02/2024	**BatchID:**	292940
Date Analyzed:	05/02/2024	**Extraction Method:**	E200.8
Instrument:	ICP-MS6	**Analytical Method:**	E200.8
Matrix:	Drinking Water	**Unit:**	µg/L
Project:	Edna Brewer Middle Sch. Drinking Water Test	**Sample ID:**	MB/LCS/LCSD-292940 2404P75-027AMS/MSD

QC Summary Report for Metals

Analyte	MB Result	MDL	RL			
Lead	ND	0.052	0.50	-	-	-

Analyte	LCS Result	LCSD Result	SPK Val	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Limit
Lead	52	52	50	104	103	85-115	0.997	20

Analyte	MS DF	MS Result	MSD Result	SPK Val	SPKRef Val	MS %REC	MSD %REC	MS/MSD Limits	RPD	RPD Limit
Lead	1	56	57	50	3.080	106	108	85-115	1.34	20

CA ELAP 1644

McCampbell Analytical, Inc.

1534 Willow Pass Rd
Pittsburg, CA 94565-1701
(925) 252-9262

CHAIN-OF-CUSTODY RECORD

WorkOrder: 2404P75 **ClientCode: GESCA** **QuoteID: 242836**

☐ WaterTrax ☐ CLIP ☐ EDF ☐ EQuIS ☒ Dry-Weight ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag
☐ Detection Summary ☐ Excel

Report to:
Walloh Gibson
Gibco Environmental Services
2163 Meeker Ave STE. 195
Richmond, CA 94804
(510) 672-9472 FAX:

Email: gibcoenvironmentalservices@gmail.com
cc/3rd Party:
PO:
Project: Edna Brewer Middle Sch. Drinking Water Test

Bill to:
Walloh Gibson
Gibco Environmental Services
2163 Meeker Ave STE. 195
Richmond, CA 94804
gibcoenvironmentalservices@gmail.com

Requested TAT: **5 days;**

Date Received: **04/30/2024**
Date Logged: **05/01/2024**

Lab ID	ClientSampID	Matrix	Collection Date	Hold	Requested Tests (See legend below) 1	2	3	4	5	6	7	8	9	10	11	12
2404P75-001	042724-1A/Flow Water-Hall Opp Counsellor Off	Water	4/27/2024 13:45	☐	A	A										
2404P75-002	042724-2B/Faucet-Sink Attend. Office	Water	4/27/2024 13:49	☐	A	A										
2404P75-003	042724-3C/Drinking Fountain Opp Rm 24	Water	4/27/2024 13:53	☐	A	A										
2404P75-004	042724-4D/Drinking Fountain-Opp. Rm 20	Water	4/27/2024 13:57	☐	A	A										
2404P75-005	042724-5E/Drinking Fountain-Hall Adj. Cafeteria	Water	4/27/2024 14:00	☐	A	A										
2404P75-006	042724-6F/Faucet-Sink-Staff Lunch Room	Water	4/27/2024 14:03	☐	A	A										
2404P75-007	042724-7G/Kitchen Prep Sink-Faucet-Sink	Water	4/27/2024 14:07	☐	A	A										
2404P75-008	042724-8G Flush/Kitchen Prep Sink-Faucet-Sink	Water	4/27/2024 14:08	☐	A	A										
2404P75-009	042724-9H/Drinking Fountain-Hall adj. RM 30	Water	4/27/2024 14:16	☐	A	A										
2404P75-010	042724-10I/Flow Water Adj. Off Bldg. 30	Water	4/27/2024 14:20	☐	A	A										
2404P75-011	042724-11J/Flow Water Adj. Off Bldg. 30	Water	4/27/2024 14:21	☐	A	A										
2404P75-012	042724-12K/Drinking Fountain-Boys Locker RM Ent.	Water	4/27/2024 14:24	☐	A	A										
2404P75-013	042724-13L/Drinking Fountain-Girls Locker RM	Water	4/27/2024 14:29	☐	A	A										
2404P75-014	042724-14M/Drinking Fountain-Girls Locker RM	Water	4/27/2024 14:32	☐	A	A										
2404P75-015	042724-15N/Exterior Drinking Fountain Gym	Water	4/27/2024 14:35	☐	A	A										

Test Legend:

1	METALSMS_DW	2	PRDisposal Fee	3		4	
5		6		7		8	
9		10		11		12	

Project Manager: Yen Cao

Prepared by: Natalie Zaragoza

Comments:

NOTE: Soil samples are discarded 60 days after receipt unless other arrangements are made (Water samples are 30 days).
Hazardous samples will be returned to client or disposed of at client expense.

McCampbell Analytical, Inc.
1534 Willow Pass Rd
Pittsburg, CA 94565-1701
(925) 252-9262

CHAIN-OF-CUSTODY RECORD

WorkOrder: 2404P75 **ClientCode: GESCA** **QuoteID: 242836**

☐ WaterTrax ☐ CLIP ☐ EDF ☐ EQuIS ☒ Dry-Weight ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag
☐ Detection Summary ☐ Excel

Report to:
Walloh Gibson
Gibco Environmental Services
2163 Meeker Ave STE. 195
Richmond, CA 94804
(510) 672-9472 FAX:

Email: gibcoenvironmentalservices@gmail.com
cc/3rd Party:
PO:
Project: Edna Brewer Middle Sch. Drinking Water Test

Bill to:
Walloh Gibson
Gibco Environmental Services
2163 Meeker Ave STE. 195
Richmond, CA 94804
gibcoenvironmentalservices@gmail.com

Requested TAT: **5 days;**

Date Received: **04/30/2024**
Date Logged: **05/01/2024**

Lab ID	ClientSampID	Matrix	Collection Date	Hold	Requested Tests (See legend below) 1	2	3	4	5	6	7	8	9	10	11	12
2404P75-016	042724-16O/Exterior Drinking Fountain-Gym	Water	4/27/2024 14:37	☐	A	A										
2404P75-017	042724-17P/Drinking Fountain-Boys Locker Room	Water	4/27/2024 14:39	☐	A	A										
2404P75-018	042724-18Q/Drinking Fountain-Hall Adj. RM 41	Water	4/27/2024 14:45	☐	A	A										
2404P75-019	042724-19R/Fauect-Sink-RM 42-Bldg 40	Water	4/27/2024 14:48	☐	A	A										
2404P75-020	042724-20S/Fauect-Sink-RM 43	Water	4/27/2024 14:52	☐	A	A										
2404P75-021	042724-21T/Faucet-Sink-RM 43	Water	4/27/2024 14:57	☐	A	A										
2404P75-022	042724-22U/Faucet-Sink-RM 43	Water	4/27/2024 14:59	☐	A	A										
2404P75-023	042724-23V/Faucet-Sink-RM 43	Water	4/27/2024 15:02	☐	A	A										
2404P75-024	042724-33W/Drinking Fountain-RM 50	Water	4/27/2024 15:08	☐	A	A										
2404P75-025	042724-34X/Faucet-Sink-Bldg 50	Water	4/27/2024 15:11	☐	A	A										
2404P75-026	042724-35Y/Faucet-Sink-Bldg 50	Water	4/27/2024 15:12	☐	A	A										
2404P75-027	042724-36Z/Drinking Fountain-RM 55	Water	4/27/2024 15:17	☐	A	A										
2404P75-028	042724-37AA/Faucet Sink-RM 55	Water	4/27/2024 15:19	☐	A	A										
2404P75-029	042724-38BB/Drinking Fountain-Rm 55	Water	4/27/2024 15:20	☐	A	A										
2404P75-030	042724-39CC/Exterior Drinking Fountain-Main Bldg	Water	4/27/2024 15:25	☐	A	A										

Test Legend:

1	METALSMS_DW	2	PRDisposal Fee	3		4	
5		6		7		8	
9		10		11		12	

Project Manager: Yen Cao

Prepared by: Natalie Zaragoza

Comments:

NOTE: Soil samples are discarded 60 days after receipt unless other arrangements are made (Water samples are 30 days).
Hazardous samples will be returned to client or disposed of at client expense.

McCampbell Analytical, Inc.
1534 Willow Pass Rd
Pittsburg, CA 94565-1701
(925) 252-9262

CHAIN-OF-CUSTODY RECORD

WorkOrder: 2404P75 **ClientCode: GESCA** **QuoteID: 242836**

☐ WaterTrax ☐ CLIP ☐ EDF ☐ EQuIS ☒ Dry-Weight ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag
☐ Detection Summary ☐ Excel

Report to:
Walloh Gibson
Gibco Environmental Services
2163 Meeker Ave STE. 195
Richmond, CA 94804
(510) 672-9472 FAX:

Email: gibcoenvironmentalservices@gmail.com
cc/3rd Party:
PO:
Project: Edna Brewer Middle Sch. Drinking Water Test

Bill to:
Walloh Gibson
Gibco Environmental Services
2163 Meeker Ave STE. 195
Richmond, CA 94804
gibcoenvironmentalservices@gmail.com

Requested TAT: **5 days;**

Date Received: **04/30/2024**
Date Logged: **05/01/2024**

Lab ID	ClientSampID	Matrix	Collection Date	Hold	Requested Tests (See legend below) 1	2	3	4	5	6	7	8	9	10	11	12
2404P75-031	042724-40DD/Exterior Drinking Fountain-Main Bldg	Water	4/27/2024 15:26	☐	A	A										
2404P75-032	042724-41EE/Exterior Drinking Fountain-Main Bldg	Water	4/27/2024 15:30	☐	A	A										
2404P75-033	042724-42FF/Exterior Drinking Fountain Bldg. 20	Water	4/27/2024 15:32	☐	A	A										
2404P75-034	042724-43GG/Exterior Drinking Fountain-Bldg 20	Water	4/27/2024 15:37	☐	A	A										
2404P75-035	042724-44HH/Exterior Drinking Fountain-Bldg 20	Water	4/27/2024 15:39	☐	A	A										
2404P75-036	042724-45II/Exterior Drinking Fountain-Bldg 20	Water	4/27/2024 15:41	☐	A	A										
2404P75-037	042724-46JJ/Exterior Drinking Fountain-Bldg 20	Water	4/27/2024 15:43	☐	A	A										
2404P75-038	042724-47KK/Flow Water Adj. Cafateria-Hallway-Musi	Water	4/27/2024 15:47	☐	A	A										
2404P75-039	042724-48LL/Flow Water Adj. Cafateria-Hallway-Musi	Water	4/27/2024 15:48	☐	A	A										

Test Legend:

1	METALSMS_DW	2	PRDisposal Fee	3		4	
5		6		7		8	
9		10		11		12	

Project Manager: Yen Cao

Prepared by: Natalie Zaragoza

Comments:

NOTE: Soil samples are discarded 60 days after receipt unless other arrangements are made (Water samples are 30 days).
Hazardous samples will be returned to client or disposed of at client expense.

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

WORK ORDER SUMMARY

Client Name: GIBCO ENVIRONMENTAL SERVICES
Client Contact: Walloh Gibson
Contact's Email: gibcoenvironmentalservices@gmail.com

Project: Edna Brewer Middle Sch. Drinking Water Test
Comments:

Work Order: 2404P75
QC Level: LEVEL 2
Date Logged: 5/1/2024

☐ WaterTrax ☐ CLIP ☐ EDF ☐ Excel ☐ EQuIS ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag

LabID	ClientSampID	Matrix	Test Name	Cont./ Comp.	Bottle & Preservative	U**	Head Space	Dry-Weight	Collection Date & Time	TAT	Test Due Date	Sediment Content	Hold	Sub Out
001A	042724-1A/Flow Water-Hall Opp Counsellor Off	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 13:45	5 days	5/7/2024	None	☐	☐
002A	042724-2B/Faucet-Sink Attend. Office	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 13:49	5 days	5/7/2024	None	☐	☐
003A	042724-3C/Drinking Fountain Opp Rm 24	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 13:53	5 days	5/7/2024	None	☐	☐
004A	042724-4D/Drinking Fountain-Opp. Rm 20	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 13:57	5 days	5/7/2024	None	☐	☐
005A	042724-5E/Drinking Fountain-Hall Adj. Cafeteria	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:00	5 days	5/7/2024	None	☐	☐
006A	042724-6F/Faucet-Sink-Staff Lunch Room	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:03	5 days	5/7/2024	None	☐	☐
007A	042724-7G/Kitchen Prep Sink-Faucet-Sink	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:07	5 days	5/7/2024	None	☐	☐
008A	042724-8G Flush/Kitchen Prep Sink-Faucet-Sink	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:08	5 days	5/7/2024	None	☐	☐

NOTES: * STLC and TCLP extractions require 2 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 3 days from sample submission).

- ISM prep requires 5 to 10 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 6 to 11 days from sample submission). Due date listed on WO summary will not accurately reflect the time needed for sample preparation.

- Organic extracts are held for 40 days before disposal; Inorganic extract are held for 30 days.

- MAI assumes that all material present in the provided sampling container is considered part of the sample - MAI does not exclude any material from the sample prior to sample preparation unless requested in writing by the client.

U = An unpreserved container was received for a method that suggests a preservation in order to extend hold time for analysis.**

Cont./Comp. = Containers /Composites

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

WORK ORDER SUMMARY

Client Name: GIBCO ENVIRONMENTAL SERVICES
Client Contact: Walloh Gibson
Contact's Email: gibcoenvironmentalservices@gmail.com

Project: Edna Brewer Middle Sch. Drinking Water Test
Comments:

Work Order: 2404P75
QC Level: LEVEL 2
Date Logged: 5/1/2024

☐ WaterTrax ☐ CLIP ☐ EDF ☐ Excel ☐ EQuIS ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag

LabID	ClientSampID	Matrix	Test Name	Cont./ Comp.	Bottle & Preservative	U**	Head Space	Dry-Weight	Collection Date & Time	TAT	Test Due Date	Sediment Content	Hold	Sub Out
009A	042724-9H/Drinking Fountain-Hall adj. RM 30	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:16	5 days	5/7/2024	None	☐	☐
010A	042724-10I/Flow Water Adj. Off Bldg. 30	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:20	5 days	5/7/2024	None	☐	☐
011A	042724-11J/Flow Water Adj. Off Bldg. 30	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:21	5 days	5/7/2024	None	☐	☐
012A	042724-12K/Drinking Fountain-Boys Locker RM Ent.	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:24	5 days	5/7/2024	None	☐	☐
013A	042724-13L/Drinking Fountain-Girls Locker RM	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:29	5 days	5/7/2024	None	☐	☐
014A	042724-14M/Drinking Fountain-Girls Locker RM	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:32	5 days	5/7/2024	None	☐	☐
015A	042724-15N/Exterior Drinking Fountain Gym	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:35	5 days	5/7/2024	None	☐	☐

NOTES: * STLC and TCLP extractions require 2 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 3 days from sample submission).

- ISM prep requires 5 to 10 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 6 to 11 days from sample submission). Due date listed on WO summary will not accurately reflect the time needed for sample preparation.

- Organic extracts are held for 40 days before disposal; Inorganic extract are held for 30 days.

- MAI assumes that all material present in the provided sampling container is considered part of the sample - MAI does not exclude any material from the sample prior to sample preparation unless requested in writing by the client.

U = An unpreserved container was received for a method that suggests a preservation in order to extend hold time for analysis.**

Cont./Comp. = Containers /Composites

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

WORK ORDER SUMMARY

Client Name: GIBCO ENVIRONMENTAL SERVICES
Client Contact: Walloh Gibson
Contact's Email: gibcoenvironmentalservices@gmail.com

Project: Edna Brewer Middle Sch. Drinking Water Test
Comments:

Work Order: 2404P75
QC Level: LEVEL 2
Date Logged: 5/1/2024

☐ WaterTrax ☐ CLIP ☐ EDF ☐ Excel ☐ EQuIS ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag

LabID	ClientSampID	Matrix	Test Name	Cont./ Comp.	Bottle & Preservative	U**	Head Space	Dry-Weight	Collection Date & Time	TAT	Test Due Date	Sediment Content	Hold	Sub Out
016A	042724-16O/Exterior Drinking Fountain-Gym	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:37	5 days	5/7/2024	None	☐	☐
017A	042724-17P/Drinking Fountain-Boys Locker Room	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:39	5 days	5/7/2024	None	☐	☐
018A	042724-18Q/Drinking Fountain-Hall Adj. RM 41	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:45	5 days	5/7/2024	None	☐	☐
019A	042724-19R/Fauect-Sink-RM 42-Bldg 40	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:48	5 days	5/7/2024	None	☐	☐
020A	042724-20S/Fauect-Sink-RM 43	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:52	5 days	5/7/2024	None	☐	☐
021A	042724-21T/Faucet-Sink-RM 43	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:57	5 days	5/7/2024	None	☐	☐
022A	042724-22U/Faucet-Sink-RM 43	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 14:59	5 days	5/7/2024	None	☐	☐
023A	042724-23V/Faucet-Sink-RM 43	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:02	5 days	5/7/2024	None	☐	☐

NOTES: * STLC and TCLP extractions require 2 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 3 days from sample submission).

- ISM prep requires 5 to 10 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 6 to 11 days from sample submission). Due date listed on WO summary will not accurately reflect the time needed for sample preparation.

- Organic extracts are held for 40 days before disposal; Inorganic extract are held for 30 days.

- MAI assumes that all material present in the provided sampling container is considered part of the sample - MAI does not exclude any material from the sample prior to sample preparation unless requested in writing by the client.

U = An unpreserved container was received for a method that suggests a preservation in order to extend hold time for analysis.**

Cont./Comp. = Containers /Composites

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

WORK ORDER SUMMARY

Client Name: GIBCO ENVIRONMENTAL SERVICES
Client Contact: Walloh Gibson
Contact's Email: gibcoenvironmentalservices@gmail.com

Project: Edna Brewer Middle Sch. Drinking Water Test
Comments:

Work Order: 2404P75
QC Level: LEVEL 2
Date Logged: 5/1/2024

☐ WaterTrax ☐ CLIP ☐ EDF ☐ Excel ☐ EQuIS ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag

LabID	ClientSampID	Matrix	Test Name	Cont./ Comp.	Bottle & Preservative	U**	Head Space	Dry-Weight	Collection Date & Time	TAT	Test Due Date	Sediment Content	Hold	Sub Out
024A	042724-33W/Drinking Fountain-RM 50	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:08	5 days	5/7/2024	None	☐	☐
025A	042724-34X/Faucet-Sink-Bldg 50	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:11	5 days	5/7/2024	None	☐	☐
026A	042724-35Y/Faucet-Sink-Bldg 50	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:12	5 days	5/7/2024	None	☐	☐
027A	042724-36Z/Drinking Fountain-RM 55	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:17	5 days	5/7/2024	None	☐	☐
028A	042724-37AA/Faucet Sink-RM 55	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:19	5 days	5/7/2024	None	☐	☐
029A	042724-38BB/Drinking Fountain-Rm 55	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:20	5 days	5/7/2024	None	☐	☐
030A	042724-39CC/Exterior Drinking Fountain-Main Bldg	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:25	5 days	5/7/2024	None	☐	☐
031A	042724-40DD/Exterior Drinking Fountain-Main Bldg	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:26	5 days	5/7/2024	None	☐	☐

NOTES: * STLC and TCLP extractions require 2 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 3 days from sample submission).

- ISM prep requires 5 to 10 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 6 to 11 days from sample submission). Due date listed on WO summary will not accurately reflect the time needed for sample preparation.

- Organic extracts are held for 40 days before disposal; Inorganic extract are held for 30 days.

- MAI assumes that all material present in the provided sampling container is considered part of the sample - MAI does not exclude any material from the sample prior to sample preparation unless requested in writing by the client.

U = An unpreserved container was received for a method that suggests a preservation in order to extend hold time for analysis.**

Cont./Comp. = Containers /Composites

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

WORK ORDER SUMMARY

Client Name: GIBCO ENVIRONMENTAL SERVICES
Client Contact: Walloh Gibson
Contact's Email: gibcoenvironmentalservices@gmail.com

Project: Edna Brewer Middle Sch. Drinking Water Test
Comments:

Work Order: 2404P75
QC Level: LEVEL 2
Date Logged: 5/1/2024

☐ WaterTrax ☐ CLIP ☐ EDF ☐ Excel ☐ EQuIS ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag

LabID	ClientSampID	Matrix	Test Name	Cont./ Comp.	Bottle & Preservative	U**	Head Space	Dry-Weight	Collection Date & Time	TAT	Test Due Date	Sediment Content	Hold	Sub Out
032A	042724-41EE/Exterior Drinking Fountain-Main Bldg	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:30	5 days	5/7/2024	None	☐	☐
033A	042724-42FF/Exterior Drinking Fountain Bldg. 20	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:32	5 days	5/7/2024	None	☐	☐
034A	042724-43GG/Exterior Drinking Fountain-Bldg 20	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:37	5 days	5/7/2024	None	☐	☐
035A	042724-44HH/Exterior Drinking Fountain-Bldg 20	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:39	5 days	5/7/2024	None	☐	☐
036A	042724-45II/Exterior Drinking Fountain-Bldg 20	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:41	5 days	5/7/2024	None	☐	☐
037A	042724-46JJ/Exterior Drinking Fountain-Bldg 20	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:43	5 days	5/7/2024	None	☐	☐

NOTES: * STLC and TCLP extractions require 2 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 3 days from sample submission).

- ISM prep requires 5 to 10 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 6 to 11 days from sample submission). Due date listed on WO summary will not accurately reflect the time needed for sample preparation.

- Organic extracts are held for 40 days before disposal; Inorganic extract are held for 30 days.

- MAI assumes that all material present in the provided sampling container is considered part of the sample - MAI does not exclude any material from the sample prior to sample preparation unless requested in writing by the client.

U = An unpreserved container was received for a method that suggests a preservation in order to extend hold time for analysis.**

Cont./Comp. = Containers /Composites

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

WORK ORDER SUMMARY

Client Name: GIBCO ENVIRONMENTAL SERVICES
Client Contact: Walloh Gibson
Contact's Email: gibcoenvironmentalservices@gmail.com

Project: Edna Brewer Middle Sch. Drinking Water Test
Comments:

Work Order: 2404P75
QC Level: LEVEL 2
Date Logged: 5/1/2024

☐ WaterTrax ☐ CLIP ☐ EDF ☐ Excel ☐ EQuIS ☐ Email ☐ HardCopy ☐ ThirdParty ☐ J-flag

LabID	ClientSampID	Matrix	Test Name	Cont./ Comp.	Bottle & Preservative	U**	Head Space	Dry-Weight	Collection Date & Time	TAT	Test Due Date	Sediment Content	Hold	Sub Out
038A	042724-47KK/Flow Water Adj. Cafateria-Hallway-Musi	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:47	5 days	5/7/2024	None	☐	☐
039A	042724-48LL/Flow Water Adj. Cafateria-Hallway-Musi	Water	E200.8 (Metals) <Lead>	1	250mL HDPE w/ HNO3	☐	☐	☐	4/27/2024 15:48	5 days	5/7/2024	None	☐	☐

NOTES: *** STLC and TCLP extractions require 2 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 3 days from sample submission).**

- ISM prep requires 5 to 10 days to complete; therefore, all TATs begin after the extraction is completed (i.e., One-day TAT yields results in 6 to 11 days from sample submission). Due date listed on WO summary will not accurately reflect the time needed for sample preparation.

- Organic extracts are held for 40 days before disposal; Inorganic extract are held for 30 days.

- MAI assumes that all material present in the provided sampling container is considered part of the sample - MAI does not exclude any material from the sample prior to sample preparation unless requested in writing by the client.

U = An unpreserved container was received for a method that suggests a preservation in order to extend hold time for analysis.**

Cont./Comp. = Containers /Composites

General COC

MAI Work Order # 2404P75

McCAMPBELL ANALYTICAL, INC.

1534 Willow Pass Rd. Pittsburg, Ca. 94565-1701

Telephone: (877) 252-9262 / Fax: (925) 252-9269

CHAIN OF CUSTODY RECORD

Turn Around Time:	1 Day Rush		2 Day Rush		3 Day Rush		STD		Quote #	242836			
J-Flag / MDL		ESL		Cleanup Approved		Dry Weight		Bottle Order #					
Delivery Format:	PDF		GeoTracker EDF		EDD		CLIP EDT (DW)		Detect Summary				

Report To: Walloh Gibson Bill To:

Company: Gibco Environmental Services

Email: gibcoenvironmentalservices@gmail.com

Alt Email: Tele: 5106729472

Project Name: Edna Brewer Middle Sch. Drinking Water Test Project #:

Project Location: 3748 13rh Ave Oakland CA 94610 PO #

Sampler Signature: Walloh D. Gibson

Analysis Requested

SAMPLE ID Location / Field Point	Sampling Date	Sampling Time	#Containers	Matrix	Preservative	EPA 200.8 (LEAD)
042724 - 1A / Flow Water - Hall Opp Counsellor Off	04-27-24	1:45 PM	1	DW	4	●
042724 - 2B / Faucet - Sink Attend. Office	04-27-24	1:49 PM	1	DW	4	●
042724 - 3C / Drinking Fountain Opp RM 24	04-27-24	1:53 PM	1	DW	4	●
042724 - 4D / Drinking Fountain - Opp. RM 20	04-27-24	1:57 PM	1	DW	4	●
042724 - 5E / Drinking Fountain - Hall Adj. Cafateria	04-27-24	2:00 PM	1	DW	4	●
042724 - 6F / Faucet - Sink - Staff Lunch Room	04-27-24	2:03 PM	1	DW	4	●
042724 - 7G / Kitchen Prep Sink - Faucet - Sink	04-27-24	2:07 PM	1	DW	4	●
042724 - 8G Flush / Kitchen Prep Sink - Faucet - Sink	04-27-24	2:08 PM	1	DW	4	●
042724 - 9H / Drinking Fountain - Hall adj. RM 30	04-27-24	2:16 PM	1	DW	4	●
042724 - 10I / Flow Water Adj. Off Bldg. 30	04-27-24	2:20 PM	1	DW	4	●

MAI clients MUST disclose any dangerous chemicals known to be present in their submitted samples in concentrations that may cause immediate harm or serious future health endangerment as a result of brief, gloved, open air, sample handling by MAI staff. Non-disclosure incurs an immediate $250 surcharge and the client is subject to full legal liability for harm suffered. Thank you for your understanding and for allowing us to work safely.

* If metals are requested for water samples and the water type (Matrix) is not specified on the chain of custody, MAI will default to metals by E200.8.

Please provide an adequate volume of sample. If the volume is not sufficient for a MS/MSD a LCS/LCSD will be prepared in its place and noted in the report.

Comments / Instructions

Relinquished By / Company Name	Date	Time	Received By / Company Name	Date	Time
Walloh D. Gibson / GIBCO ENVIRONMENTAL	04-29-24	3:00PM	[illegible] / MAI	4/30/24	15
[illegible] / MAI	4/30/24	1450	[illegible]	4/30/24	1450

Matrix Code: DW=Drinking Water, GW=Ground Water, WW=Waste Water, SW=Seawater, S=Soil, SL=Sludge, A=Air, WP=Wipe, O=Other

Preservative Code: 1=4°C 2=HCl 3=H_2SO_4 4=HNO_3 5=NaOH 6=ZnOAc/NaOH 7=None

Temp 0.4 °C Initials wt

MAI Work Order #_____

McCAMPBELL ANALYTICAL, INC.

1534 Willow Pass Rd. Pittsburg, Ca. 94565-1701
Telephone: (877) 252-9262 / Fax: (925) 252-9269

CHAIN OF CUSTODY RECORD

Turn Around Time:	1 Day Rush		2 Day Rush		3 Day Rush		STD		Quote #	242836			
J-Flag / MDL		ESL		Cleanup Approved		Dry Weight		Bottle Order #					
Delivery Format:	PDF		GeoTracker EDF		EDD		CLIP EDT (DW)		Detect Summary				

Report To: Walloh Gibson Bill To:

Company: Gibco Environmental Services

Email: gibcoenvironmentalservices@gmail.com

Alt Email: Tele: 5106729472

Project Name: Edna Brewer Middle Sch. Drinking Water Test Project #:

Project Location: 3748 13rh Ave Oakland CA 94610 PO #

Sampler Signature: Walloh A. GIBSON

Analysis Requested

SAMPLE ID Location / Field Point	Sampling Date	Sampling Time	#Containers	Matrix	Preservative	EPA 200.8 (LEAD)
042724 - 11J / Flow Water Adj. Off Bldg. 30	04-27-24	2:21 PM	1	DW	4	●
042724 - 12K / Drinking Fountain - Boys Locker RM Ent.	04-27-24	2:24 PM	1	DW	4	●
042724 - 13L / Drinking Fountain - Girls Locker RM Ent	04-27-24	2:29 PM	1	DW	4	●
042724 - 14M / Drinking Fountain - Girls Locker RM	04-27-24	2:32 PM	1	DW	4	●
042724 - 15N / Exterior Drinking Fountain Gym	04-27-24	2:35 PM	1	DW	4	●
042724 - 16O / Exterior Drinking Fountain - Gym	04-27-24	2:37 PM	1	DW	4	●
042724 - 17P / Drinking Fountain - Boys Locker RM	04-27-24	2:39 PM	1	DW	4	●
042724 - 18Q / Drinking Fountain - Hall Adj. RM 41	04-27-24	2:45 PM	1	DW	4	●
042724 - 19R / Faucet - Sink - RM 42 - Bldg 40	04-27-24	2:48 PM	1	DW	4	●
042724 - 20S / Faucet - Sink - RM 43	04-27-24	2:52 PM	1	DW	4	●

MAI clients MUST disclose any dangerous chemicals known to be present in their submitted samples in concentrations that may cause immediate harm or serious future health endangerment as a result of brief, gloved, open air, sample handling by MAI staff. Non-disclosure incurs an immediate $250 surcharge and the client is subject to full legal liability for harm suffered. Thank you for your understanding and for allowing us to work safely.

* If metals are requested for water samples and the water type (Matrix) is not specified on the chain of custody, MAI will default to metals by E200.8.

Please provide an adequate volume of sample. If the volume is not sufficient for a MS/MSD a LCS/LCSD will be prepared in its place and noted in the report.

Comments / Instructions

Relinquished By / Company Name	Date	Time	Received By / Company Name	Date	Time
Walloh A. GIBSON / GIBCO ENVIRONMENTAL	04-29-24	3:00PM	MAI	4/30/24	18
MAI	4/30/24	1450	Natalie Coleman	4/30/24	1450

Matrix Code: DW=Drinking Water, GW=Ground Water, WW=Waste Water, SW=Seawater, S=Soil, SL=Sludge, A=Air, WP=Wipe, O=Other

Preservative Code: 1=4°C 2=HCl 3=H_2SO_4 4=HNO_3 5=NaOH 6=ZnOAc/NaOH 7=None

Temp _____ °C Initials _____

MAI Work Order #_____

McCAMPBELL ANALYTICAL, INC.

1534 Willow Pass Rd. Pittsburg, Ca. 94565-1701
Telephone: (877) 252-9262 / Fax: (925) 252-9269

CHAIN OF CUSTODY RECORD

Turn Around Time:	1 Day Rush		2 Day Rush		3 Day Rush		STD		Quote #	242836	
J-Flag / MDL		ESL		Cleanup Approved		Dry Weight		Bottle Order #			
Delivery Format:	PDF		GeoTracker EDF		EDD		CLIP EDT (DW)		Detect Summary		

Report To: Walloh Gibson Bill To:

Company: Gibco Environmental Services

Email: gibcoenvironmentalservices@gmail.com

Alt Email: Tele: 5106729472

Project Name: Edna Brewer Middle Sch. Drinking Water Test Project #:

Project Location: 3748 13rh Ave Oakland CA 94610 PO #

Sampler Signature: Walloh A. Gibson

Analysis Requested

SAMPLE ID Location / Field Point	Sampling Date	Sampling Time	#Containers	Matrix	Preservative	EPA 200.8 (LEAD)
042724 - 21T / Faucet - Sink - RM 43	04-27-24	2:57 PM	1	DW	4	●
042724 - 22U / Faucet - Sink - RM 43	04-27-24	2:59 PM	1	DW	4	●
042724 - 23V / Faucet - Sink - RM 43	04-27-24	3:02 PM	1	DW	4	●
042724 - 33W / Drinking Fountain - RM 50	04-27-24	3:08 PM	1	DW	4	●
042724 - 34X / Faucet - Sink - Bldg 50	04-27-24	3:11 PM	1	DW	4	●
042724 - 35Y / Faucet - Sink - Bldg 50	04-27-24	3:12 PM	1	DW	4	●
042724 - 36Z / Drinking Fountain - RM 55	04-27-24	3:17 PM	1	DW	4	●
042724 - 37AA / Faucet Sink - RM 55	04-27-24	3:19 PM	1	DW	4	●
042724 - 38BB / Drinking Fountain -RM 55	04-27-24	3:20 PM	1	DW	4	●
042724 - 39CC / Exterior Drinking Fountain - Main Bldg Opp Bldg 50	04-27-24	3:25 PM	1	DW	4	●

MAI clients MUST disclose any dangerous chemicals known to be present in their submitted samples in concentrations that may cause immediate harm or serious future health endangerment as a result of brief, gloved, open air, sample handling by MAI staff. Non-disclosure incurs an immediate $250 surcharge and the client is subject to full legal liability for harm suffered. Thank you for your understanding and for allowing us to work safely.

* If metals are requested for water samples and the water type (Matrix) is not specified on the chain of custody, MAI will default to metals by E200.8.

Please provide an adequate volume of sample. If the volume is not sufficient for a MS/MSD a LCS/LCSD will be prepared in its place and noted in the report.

Comments / Instructions

Relinquished By / Company Name	Date	Time	Received By / Company Name	Date	Time
Walloh A. Gibson / GIBCO ENVIRONMENTAL	04-29-24	3:00 PM	[illegible]	4/30/24	10
[illegible]	4/30/24	1450	[illegible]	4/30/24	1450

Matrix Code: DW=Drinking Water, GW=Ground Water, WW=Waste Water, SW=Seawater, S=Soil, SL=Sludge, A=Air, WP=Wipe, O=Other

Preservative Code: 1=4°C 2=HCl 3=H_2SO_4 4=HNO_3 5=NaOH 6=ZnOAc/NaOH 7=None

Temp _____ °C Initials _____

MAI Work Order #_____

McCAMPBELL ANALYTICAL, INC.

1534 Willow Pass Rd. Pittsburg, Ca. 94565-1701
Telephone: (877) 252-9262 / Fax: (925) 252-9269

CHAIN OF CUSTODY RECORD

Turn Around Time:	1 Day Rush		2 Day Rush		3 Day Rush		STD		Quote #	242836			
J-Flag / MDL		ESL		Cleanup Approved		Dry Weight		Bottle Order #					
Delivery Format:	PDF		GeoTracker EDF		EDD		CLIP EDT (DW)		Detect Summary				

Report To: Walloh Gibson Bill To:

Company: Gibco Environmental Services

Email: gibcoenvironmentalservices@gmail.com

Alt Email: Tele: 5106729472

Project Name: Edna Brewer Middle Sch. Drinking Water Test Project #:

Project Location: 3748 13rh Ave Oakland CA 94610 PO #

Sampler Signature: Walloh D. Gibson

Analysis Requested

SAMPLE ID Location / Field Point	Sampling Date	Sampling Time	#Containers	Matrix	Preservative	EPA 200.8 (LEAD)
042724 - 40DD / Exterior Drinking Fountain - Main Bldg Opp Bldg 50	04-27-24	3:26 PM	1	DW	4	●
042724 - 41EE / Exterior Drinking Fountain - Main Bldg Opp Bldg 50	04-27-24	3:30 PM	1	DW	4	●
042724 - 42FF / Exterior Drinking Fountain - Main Bldg Opp Bldg 50	04-27-24	3:32 PM	1	DW	4	●
042724 - 43GG / Exterior Drinking Fountain - Bldg. 20	04-27-24	3:37 PM	1	DW	4	●
042724 - 44HH / Exterior Drinking Fountain - Bldg 20	04-27-24	3:39 PM	1	DW	4	●
042724 - 45II - Exterior Drinking Fountain - Bldg 20	04-27-24	3:41 PM	1	DW	4	●
042724 - 46JJ / Exterior Drinking Fountain - Bldg. 20	04-27-24	3:43 PM	1	DW	4	●
042724 - 47KK / Flow Water Adj. Cafateria - Hallway - Music RM	04-27-24	3:47 PM	1	DW	4	●
042724 - 48LL / Flow Water Adj. Cafateria - Hallway - Music RM	04-27-24	3:48 PM	1	DW	4	●

MAI clients MUST disclose any dangerous chemicals known to be present in their submitted samples in concentrations that may cause immediate harm or serious future health endangerment as a result of brief, gloved, open air, sample handling by MAI staff. Non-disclosure incurs an immediate $250 surcharge and the client is subject to full legal liability for harm suffered. Thank you for your understanding and for allowing us to work safely.

* If metals are requested for water samples and the water type (Matrix) is not specified on the chain of custody, MAI will default to metals by E200.8.

Please provide an adequate volume of sample. If the volume is not sufficient for a MS/MSD a LCS/LCSD will be prepared in its place and noted in the report.

Comments / Instructions

Relinquished By / Company Name	Date	Time	Received By / Company Name	Date	Time
Walloh D Gibson / GIBCO ENVIRONMENTAL	04-29-24	3:00PM	[illegible] / MAI	4/30/24	10
[illegible] / MAI	4/30/24	1450	[illegible]	4/30/24	1450

Matrix Code: DW=Drinking Water, GW=Ground Water, WW=Waste Water, SW=Seawater, S=Soil, SL=Sludge, A=Air, WP=Wipe, O=Other

Preservative Code: 1=4°C 2=HCl 3=H_2SO_4 4=HNO_3 5=NaOH 6=ZnOAc/NaOH 7=None

Temp _____ °C Initials _____

McCampbell Analytical, Inc.
"When Quality Counts"

1534 Willow Pass Road, Pittsburg, CA 94565-1701
Toll Free Telephone: (877) 252-9262 / Fax: (925) 252-9269
http://www.mccampbell.com / E-mail: main@mccampbell.com

Sample Receipt Checklist

Client Name:	Gibco Environmental Services	Date and Time Received:	4/30/2024 14:50
Project:	Edna Brewer Middle Sch. Drinking Water Test	Date Logged:	5/1/2024
		Received by:	Natalie Zaragoza
WorkOrder №:	2404P75 Matrix:	Logged by:	Natalie Zaragoza
Carrier:	Antonio Mason (MAI Courier)		

Chain of Custody (COC) Information

Chain of custody present?	Yes ☑	No ☐	
Chain of custody signed when relinquished and received?	Yes ☑	No ☐	
Chain of custody agrees with sample labels?	Yes ☑	No ☐	
Sample IDs noted by Client on COC?	Yes ☑	No ☐	
Date and Time of collection noted by Client on COC?	Yes ☑	No ☐	
Sampler's name noted on COC?	Yes ☑	No ☐	
COC agrees with Quote?	Yes ☑	No ☐	NA ☐

Sample Receipt Information

Custody seals intact on shipping container/cooler?	Yes ☐	No ☐	NA ☑
Custody seals intact on sample bottles?	Yes ☐	No ☐	NA ☑
Shipping container/cooler in good condition?	Yes ☑	No ☐	
Samples in proper containers/bottles?	Yes ☑	No ☐	
Sample containers intact?	Yes ☑	No ☐	
Sufficient sample volume for indicated test?	Yes ☑	No ☐	

Sample Preservation and Hold Time (HT) Information

All samples received within holding time?	Yes ☑	No ☐	NA ☐
Samples Received on Ice?	Yes ☑	No ☐	
(Ice Type: WET ICE)			
Sample/Temp Blank temperature	Temp: 0.4°C		NA ☐
ZHS conditional analyses: VOA meets zero headspace requirement (VOCs, TPHg/BTEX, RSK)?	Yes ☐	No ☐	NA ☑
Sample labels checked for correct preservation?	Yes ☑	No ☐	
pH acceptable upon receipt (Metal: <2; Nitrate 353.2/4500NO3: <2; 522: <4; 218.7: >8)?	Yes ☑	No ☐	NA ☐
UCMR Samples:			
pH tested and acceptable upon receipt (200.7: ≤2; 533: 6 - 8; 537.1: 6 - 8)?	Yes ☐	No ☐	NA ☑
Free Chlorine tested and acceptable upon receipt (<0.1mg/L) [not applicable to 200.7]?	Yes ☐	No ☐	NA ☑

Comments: